

Certificate of Analysis

Spectrophotometric Chlorophyll Standards

22 July 2019

Lot CAS162-01

Original Chlorophyll Solution A:

Mixed: 17 July 2019

Spectrophotometric Data: 17 July 2019

Abs₆₆₄ = 0.94344 Abs₇₅₀ = 4.6631 E-02

Chlorophyll concentration determination (Jeffrey Method, 1997)

Chlorophyll-a = 10.23 mg/L

Chlorophyll dilution A is provided with this note. The concentration value is guaranteed through July, 2020, but degradation depends on the handling of the standard. Standards must be kept at temperatures below -20° C and in the dark until use. Spectrophotometers should be standardized in subdued light with this solution warmed to room temperature, about 22° C. The determination of chlorophyll-a in aqueous acetone solvent is a sensitive technique. Please pay special attention to the configuration of the spectrophotometer itself. A bandpass of 2.0 nm or less is necessary to properly use this standard.



John Salinas
Analytical Chemist

22 July 2019
Date



Reliable Instruments for an Unreliable World
1995 N. 1st Street Telephone 408-749-0994
San Jose, CA 95112 Toll Free 877-316-8049
www.turnerdesigns.com Fax 408-749-0998