

Certificate of Analysis

Spectrophotometric Chlorophyll Standards

20 January 2020

Lot CAS163-03

Original Chlorophyll Solution A:

Mixed: 22 October 2019

Spectrophotometric Data: 14 January 2020

Abs₆₆₄ = 0.89127 Abs₇₅₀ = -2.4414 E-03

Chlorophyll concentration determination (Jeffrey Method, 1997)

Chlorophyll-a = 10.19 mg/L

Chlorophyll dilution A is provided with this note. The concentration value is guaranteed through January 2021, but degradation depends on the handling of the standard. Standards must be kept at temperatures below -20° C and in the dark until use. Spectrophotometers should be standardized in subdued light with this solution warmed to room temperature, about 22° C. The determination of chlorophyll-a in aqueous acetone solvent is a sensitive technique. Please pay special attention to the configuration of the spectrophotometer itself. A bandpass of 2.0 nm or less is necessary to properly use this standard.



20 January 2020

John Salinas
Analytical Chemist

Date



Reliable Instruments for an Unreliable World
1995 N. 1st Street Telephone 408-749-0994
San Jose, CA 95112 Toll Free 877-316-8049
www.turnerdesigns.com Fax 408-749-0998